

THAPAR INSTITUTE

OF ENGINEERING & TECHNOLOGY

(Deemed-to-be-University u/s 3 of the UGC Act, 1956)

Thapar Technology Campus, Bhadson Road

Patiala 147 004 Punjab India

Phone : +91-175-2393917 (O), 2393870, 3869

Email : npsingh@thapar.edu, atul.gupta@thapar.edu

URL : www.thapar.edu

Enquiry No.TIET/CS/AG/18-19/18018

Date: April 11, 2018

Sub: Request for Quotation(s) for the supply of Analytical Weight Balance

Dear Sir

We shall be grateful if you kindly let us have your lowest quotations for the following materials. THE QUOTATIONS SHOULD REACH THE UNDERSIGNED LATEST BY 18.4.2018 through courier or e-mail accompanied by appropriate illustrative literature/catalogues/pamphlets/technical details, samples and specifications as the case may be. On the quotation envelope/ subject the Enquiry Number & Date should be mentioned on the top of the Envelope/mail subject.

Sr. No.	Item Name	Qty.
1.	Analytical Weight Balance Specificatoin Sheet Attached:	1 Nos.

The offer sent by you must furnish the following details:

1. Name, Make & specifications of each item.
2. Cost of the item with MRP.(Treat it mandatory)
3. Educational discount if any.
4. Validity of quotation should be at least 60 Days.
5. GST . %
6. Delivery FOR Thapar University, Patiala
7. Insurance, Freight & other charges if any.
8. Minimum Delivery Period.
9. Payment terms. Net 30 days against delivery & satisfactory installation at Thapar Inst of Engg & Tech.
10. Guarantee / Warranty Information.
11. Also please share your Companies Turnover and Market Share along with the ofer.

Regards,

Sd/-

Head Commercial

Analytical Weight Balance

The analytical weight balance is required as per the following Specifications:

- (1) Capacity: 220 gm
- (2) With Draft Shield
- (3) Balance Accuracy: as per the following details
 - (i) Accuracy of 0.1 mg (for specimen Mass between 5 gm and 50 gm)
 - (ii) Accuracy of 1 mg (for specimen Mass > 50 gm)