



**THAPAR INSTITUTE**  
OF ENGINEERING & TECHNOLOGY  
(Deemed to be University)

Thapar Technology Campus, Bhadson Road

Patiala-147004, Punjab India

Phone: +91 175 2393869,70

Email: [pushap.raj@thapar.edu](mailto:pushap.raj@thapar.edu); [harjee.banga@thapar.edu](mailto:harjee.banga@thapar.edu)

URL: Thapar.edu

**Enquiry No. TIET/CS/HB/18-19/18420**

**Date : December 11, 2018**

**Sub: Request for Quotation(s) for the Supply of Hot Chamber Oil Stirrer Bath**

Dear Sir

We shall be grateful if you kindly let us have your lowest **quotations** for the following materials. THE QUOTATIONS SHOULD REACH THE UNDERSIGNED **LATEST BY 18-12-2018** through courier or e-mail accompanied by appropriate illustrative literature/catalogues/pamphlets/technical details, samples and specifications as the case may be. **On the quotation envelope/ subject the Enquiry Number & Date should be mentioned on the top of the Envelope/mail subject.**

Sr. No.	Item Name	Qty.
1.	Hot Chamber Oil Stirrer Bath  (Specifications Attached)	1 No.

**The offer sent by you must furnish the following details:**

1. Name, Make & specifications of each item.
2. Price Breakup itemwise with MRP.  
(Treat it mandatory)
3. Educational discount if any.
4. Validity of quotation should be at least 60 Days.
5. GST %
6. Delivery FOR Central Stores TIET. Patiala
7. Insurance, Freight & other charges if any.
8. Minimum Delivery Period.
9. Payment terms. Net 60 days against delivery & satisfactory installation at Thapar Institute, Patiala
10. Guarantee / Warranty Information.
11. Also please share your Companies Turnover and Market Share along with the offer.

Regards,

**Head Commercial**

## SPECIFICATIONS

### Hot Chamber Oil Stirrer Bath:

1. Temperature range must be from 50°C to 250°C.
2. Shape of the bath must be rectangular.
3. Capacity of the oil bath must be atleast 30 liters
4.  $\pm 1^{\circ}\text{C}$  Accuracy required for oil bath.
5. Electronic digital temperature controller required for controlling temperature.
6. High speed stirrer must be installed for internal oil circulation.
7. External mechanical stirrer must be installaed for mixing the samples kept in oil bath.
8. Display of bath must be digital.