

Thapar Technology Campus, Bhadson Road
Patiala-147004, Punjab India
Phone: +91 175 2393869, 70
Email: npsingh@thapar.edu; harjee.banga@thapar.edu
URL: Thapar.edu

Enquiry No. TIET/CS/HB/21-22/21378
Date : August 9, 2021

Sub: Request for Quotation(s) for the **d₃₃ Meter**

Dear Sir

We shall be grateful if you kindly let us have your lowest **quotations** for the following materials. THE QUOTATIONS SHOULD REACH THE UNDERSIGNED **LATEST BY August 17, 2021 @ 11:00 Am** through **courier or e-mail** accompanied by appropriate illustrative literature/catalogues/pamphlets/technical details, samples and specifications as the case may be. **On the quotation envelope/ subject the Enquiry Number & Date should be mentioned on the top of the Envelope/mail subject.**

Sr. No.	Item Name	Qty.
1.	d₃₃ Meter (Specifications attached)	1

The offer sent by you must furnish the following details:

- | | |
|---|--|
| 1. Name, Make & specifications of each item. | 6. Delivery FOR Central Stores TIET, Patiala |
| 2. Price Breakup itemwise with MRP. (Treat it mandatory) | 7. Insurance, Freight & other charges if any. |
| 3. Educational discount if any. | 8. Minimum Delivery Period. |
| 4. Validity of quotation should be at least 60 Days. | 9. Payment terms. Net 30 days against delivery & satisfactory installation at Thapar Institute, Patiala against 10% PBG as per TIET format. |
| 5. GST % | 10. Guarantee / Warranty Information. |
| | 11. Also please share your Companies Turnover and Market Share along with the offer. |

Regards,

Head Commercial

SPECIFICATIONS

Technical specifications for d₃₃ meter

Sr. No.	Parameters	Specifications	Comply (Y/N)	Remarks if No
1.	d ₃₃	1 to 2000 pC/N		
2.	Accuracy	2%		
3.	Resolution	0.1 pC/N		
4.	Force	0.25 N		
5.	Frequency Range	10-100 Hz		
6.	Polarity indication	Indicates polarity on upper face of test element in compression		
7.	Shunt Capacitance	0.1 pF		
8.	Power	220V, 50/60 Hz, 20W		
9.	Operable temperature range	0-50 °C		
10.	Warranty	1 Year		